

CLASSIC SERIES | Cortical™




TISSUE LEVEL IMPLANT


BONE TYPES	All Bone Types Recommended for Soft Bone Type
PROSTHETICS PLATFORM	Internal hex
DESIGN FEATURES	<ul style="list-style-type: none"> • Large cutting surface area • Tapered thread and tapered core body • Wide and sharp threads • Smooth machine surface basket neck
CLINICAL BENEFITS	<ul style="list-style-type: none"> • Self tapping • High primary stability • Minimal drilling • Immediate loading - excellent solution for implantation in extractions sites

ORDERING INFORMATION

Ø D (mm)	D0	L (mm)	L1 (mm)*	Ref. No
4.0	1.8	10	6	NM-M4010
		11.5	7.5	NM-M4011
		13	9	NM-M4013
		16	12	NM-M4016
		18	14	NM-M4018
		20	16	NM-M4020
5.0	1.8	10	6	NM-M5010
		11.5	7.5	NM-M5011
		13	9	NM-M5013
		16	12	NM-M5016
6.0	2.0	10	6	NM-M6010
		11.5	7.5	NM-M6011
		13	9	NM-M6013
		16	12	NM-M6016


Cover Screw Included with all implants  NM-S5023

* L1 - RBM Surface Treated length to place within the osteotomy.






CLASSIC SERIES | Cortical™

RECOMMENDED STRAIGHT DRILL PROTOCOL



RECOMMENDED STEP DRILL PROTOCOL



	Drill to mark osteotomy site		Drill osteotomy to implant		Drill osteotomy partially according to implant
-------------------------------------------------------------------------------------	------------------------------	-------------------------------------------------------------------------------------	----------------------------	-------------------------------------------------------------------------------------	------------------------------------------------

The recommended drill protocol procedure should not replace the dentist's/surgeon's judgment. The implants may be loaded for immediate function when good primary stability (above 35 Ncm) has been achieved and with appropriate occlusal loading.